



⑪ Publication number: **0 408 483 A1**

⑫ **EUROPEAN PATENT APPLICATION**

②① Application number: **90500063.4**

⑤① Int. Cl.⁵: **A61M 5/172**

②② Date of filing: **03.07.90**

The title of the invention has been amended
(Guidelines for Examination in the EPO, A-III,
7.3).

③③ Priority: **04.07.89 ES 8902365**

④③ Date of publication of application:
16.01.91 Bulletin 91/03

⑥④ Designated Contracting States:
AT BE CH DE DK FR GB GR IT LI LU NL SE


⑦① Applicant: **R.G.B. MEDICAL DEVICES S.A.**
Basilica, 20
E-28020 Madrid(ES)

⑦② Inventor: **Borches Jacassa, Daniel**
Caleruega, 31
E-28033 Madrid(ES)
Inventor: **Ruiz Fernandez, Ricardo**
Eduardo Dato, 3
E-28010 Madrid(ES)

⑦④ Representative: **Gonzalez Vacas, Eleuterio**
Calle Sagasta, 4
E-28004 Madrid(ES)

⑤④ **Drug delivery system.**

⑤⑦ The object of the invention is to improve methods for delivering vasoactive drugs by suitable infusion devices, controlled by a computerized device which controls and adjusts automatically in a closed loop, establishing the dosage in order to maintain blood pressure within a predefined set of values. These improvements involve using equipment which automatically controls and adapts the dosage for the different vasoactive medicine infusion pumps. This equipment is comprised of a controller (1), a monitor (2) and a high precision infusion pump (3), while the main part of this equipment is the controller (1) which includes programmed logic units to carry out this function, as well as the necessary input and control devices.



EP 0 408 483 A1

IMPROVEMENTS TO DRUG DELIVERY SYSTEMS

OBJECT OF THE INVENTION

As set out in the heading of this descriptive report, the present invention refers to an improved method for delivering vasoactive drugs using infusion equipment with a computerized device which controls and
 5 adjusts dosage automatically in a closed loop and is used to control blood pressure. This improvement is based on including a unit which automatically controls and adapts the dosage for different vasoactive medicine infusion pumps.

10 HISTORY OF THE INVENTION

Parenteral use of vasoactive medicines (nitrites and nitrates) is now common practice for the clinical treatment of blood pressure and other cardiovascular pathologies (myocardial infarction, angina, serious congestive heart failure, heart surgery, etc.). They are given intravenously, by continuous infusion, in most
 15 cases. For some medicines the dosage is of the dosage-response kind and the infusion pattern is usually different for each patient. Even for the same patient this pattern varies with time. In these cases, the "manual" dosage is usually given by starting off with a very low dose which is gradually increased until the desired response is obtained. Once the objective has been achieved, continuous supervision is required to maintain the same response levels as well as constant modifications of the dosage by the staff in charge of
 20 this control.

In the case of fast acting drugs whose effect wears off quickly, special characteristics are involved with regard to dosage as well as to control dosage response. This process requires continuous monitoring with close supervision. This "manual" control of dosage and response, is very time consuming and therefore involves diverting attention from other equally important tasks.

25 Different strategies have been considered traditionally in order to develop a safe and effective method of controlling dosage and response automatically. However, equipment which is sufficiently satisfactory for clinical purposes, does not exist as yet.

Ideally, a vasoactive drug infusion controller unit for blood pressure control, should achieve sustained hypotension, within a set time while maintaining it within close limits and, should react quickly and correctly
 30 when faced with any circumstances which attempt to upset the balanced achieved.

Traditionally, algorithms for automatic control of physiological variables have been designed based on mathematical formulas which establish a direct and unique relationship between input variables (present and past data concerning blood pressure) and the output variable (dosage to be infused).

One of the most widely used control algorithms is the so called PID (Proportional, Integrator, Derivative)
 35 controller, by means of which the amount of drug infused is a preestablished linear combination of the error between present and desired output, its integral and derivative. In spite of the fact that this kind of control is easy to implement, in practice it has not been found satisfactory due to the fact that there is no preestablished method of selecting the relative weights of the three linear combination terms, when the description of the physiological system is unknown.

40 Another type of control is based on what are known as optimum algorithms, perhaps the most representative example of which is the so called "generalized minimum variance" one.

Optimum controllers assume that the system is not known and they calculate the control required from two aspects: 1) the system's estimated characteristics (for which a pattern is made beforehand); and 2) operating criteria which should be minimized (error, amount of drug infused, etc.).

45 The advantage of using this type of controller resides in the fact that very sophisticated calculation techniques exist; their benefits can be predicted and are "optimum" in the sense that they optimize a certain criterion.

As a disadvantage we can point out the fact that while they are capable of operating best when the system is represented precisely, control can be inadequate when the system is represented imprecisely. To
 50 represent the system precisely, the optimum controller requires permanent stimulation to obtain the characteristic parameters. Paradoxically, when the desired conditions are achieved, stimulation must be at a minimum (to maintain these) and therefore both the source of information and control is consequently lost.

Other controllers which have mostly failed to provide suitable benefits are those which tend to cancel the physiological system's poles. This is not feasible in practice, since an infinite number of poles are involved in characterizing the physiological system completely, and they vary with time based on multiple

factors, the majority of which are random.

Taking into account also that physiological systems are also not well characterized (different sensitivity of patients to the same medicine, changes in the same patient's sensitivity to any drug in time (for example, tachyphylaxis, etc.)), it is desirable to use controllers which either do not depend on the cardiovascular
 5 system to be controlled or, if they do so, adapt automatically to its changes.

DESCRIPTION OF THE INVENTION

10 The benefit of the controller proposed by the invention, which is based on improving methods of delivering vasoactive drugs, is based on implementing a new control concept based on qualitative characterization of the variables to be controlled. This controller introduces logical analysis to describe those "very complex" or "badly defined" systems which do not accept precise mathematical analysis. Its main features are as follows: 1) it uses qualitative (linguistic) characterization of the control phases rather
 15 than quantitative variables (numerical ones); and 2) it establishes the relationships between variables by means of conditional statements.

This system "based on rules" is a means of treating with "imprecision", modelling the process of logical human thinking, in order to control systems which cannot be modelled exactly. The main difference between this approach to the problem and conventional techniques is that the system "based on rules"
 20 uses qualitative information, while the latter require rigid analytical relationships to describe the process.

This system attempts to simulate the process of human thinking, so it starts from the premise that human beings "categorize" information (including numerical values), with linguistic labels. For example, the blood pressure descent gradient can be classified as: "very quick", "quick", "acceptable", "slow", "very slow", etc.

25 The limits of these labels are not precisely defined numerically, but permit overlapping in such a way that, in accordance with other conditioning factors, any numerical value can belong to different labels.

The rules of Bool's algebra allow labels or groups of labels to be combined in order to establish the present condition from among the controller's range of possibilities and to take the relevant steps, based on this.

30 There are no standard rules for calibrating a control system based on rules such as the one described and, both prior experience as well as lots of experiments are required in order to identify the parameters used by the operator, deduce the linguistic labels used by the operator to classify the values measured for each parameter, and if possible, identify the range of each label and decide on the conditional rules which relate these linguistic labels to special actions of control.

35 This kind of control criterion is implemented by computerized equipment which by means of a closed loop system, compiles the analogical or digital information originating from a patient's monitoring (from conventional monitoring equipment, or from a transducer) and which, once the information has been processed, determined and sends dosage orders to an infusion pump.

The equipment has three basic units: a controller, a pressure monitor and an infusion pump with the
 40 possibility of combining these three devices in several ways: controller (connected to a conventional monitor and to an infusion pump of the kind sold at present), controller plus pressure monitor (connected to an infusion pump of the kind sold at present) or a controller plus a pressure monitor plus an infusion pump in the same unit.

In any case the infusion pump must be a high precision one and the pressure monitor must have a
 45 signal amplifying module with isolation for the blood pressure signal.

The main part of the equipment is the controller which includes a 256 K ROM memory, 64 K RAM microcontroller, visualizer which can generate its own characters, three standard RS-232 series communication outputs, real time clock, keyboard, indicators, microprocessor system back-up logic, battery, battery
 50 charging system, battery level detector and plate input level generator.

Among the indicators are included the on/off, battery level, running control and infusion order indicators. With regard to the microprocessor system's backup logic we can point out decodification, reset generator, pressure reference, latch, etc. The power supply board includes a DC-DC convertor, transformer with insulation, rectifier, regulator and mains-battery switch.

DESCRIPTION OF THE DRAWINGS

To complete the description being made and to help the characteristics of the invention to be better

understood, a single sheet of drawings is attached to the present descriptive report, as an integral part of same, showing in a single figure, the configuration which permits improvements to methods for delivering vasoactive drugs by infusion pumps by means of a computerized device which automatically controls and adjusts dosage in order to control blood pressure. This is the most widely used configuration, from which the specific configurations already described in this descriptive report are derived.

BEST WAY OF MAKING THE INVENTION

A look at the drawing shows us that the improvement to methods for delivering vasoactive drugs claimed, is based on using a controller (1), which has a monitor (2) and an infusion pump (3). The three items can comprise a single unit or optionally, a commercial monitor (2) or monitor (2) and infusion pump (3) can be used, since the special characteristics of the invention are in the controller (1) to be precise.

As we have already said, the controller (1) is organized on the basis of a ROM AND RAM memory, microcontroller, a visualizer which is capable of generating its own characters, three standard series RS-232 communication outputs, a real time clock, keyboard, on/off indicators, as well as battery level, running control and infusion order indicators, as well as the microprocessor system backup logic. The microcontroller on its part will be organized based on a central processing unit, a series communication device, two parallel input/output communication doors of eight bits each, and eight analogical/ digital conversion channels. With regard to the microprocessor system backup logic, besides the decodifying devices, there is a reset generator, a pressure reference, a latch, etc.

It is also obvious that the controller will be assisted by a supply battery, with its corresponding charging system, a level detector and a power supply board. A DC-DC convertor will be included in the latter, as well as a transformer with insulation, the corresponding rectification and regulation circuits, as well as a mains-battery switch.

With regard to the monitor (2), it must include a signal amplification module with isolation for intraarterial pressure signal.

Finally, the infusion pump (3) must be a high precision one.

As a means of further explanation of the EQUIPMENT above described, and in order to provide a better understanding of the subject of the present patent, the hardware modules are indicated as well as unique or most significant operating features of the equipment. (As explained further abstracted in the claims 1,2). any longer in order for any expert on the matter to understand the implications of the invention and the advantages derived from same.

The materials, form, size and arrangement of the items may be varied provided that this does not involve altering the essential part of the invention.

The terms in which this report has been written should always be taken in the widest possible sense and are not restrictive.

HARDWARE DESCRIPTION OF EQUIPMENT

The equipment contains the following operating functional modules:

- 1) Microprocessor system.
- 2) Analog pressure signal adapter.
- 3) Switching Power supply.
- 4) Battery charger plus battery backup system.
- 5) Graphic Digital Display.
- 6) keyboard

Next, the most important electronic component are presented, for each different functional module:

MODULE 1: MICROPROCESSOR SYSTEM

Circuit components of module 1 are listed:

- a) Microcontroller 68HC11, Motorola;
- b) Three standard RS232c serial communication ;
- c) Up to kbytes of RAM memory ;
- d) Up to 256 kbytes of EPROM memory ;

- e) 32 digital TTL compatible input-output;
- f) 8 Analog/Digital converter;
- g) Acoustic signal adapter;
- h) Display Backlight Actuator ;
- 5 i) Alarm visual indicator;
- j) keyboard controller;

MODULE 2: ANALOG PRESSURE SIGNAL ADAPTER

10

Circuit components of module 2 are listed. The input signal comes from the arterial blood pressure transducer.

- a) Differential Amplifier with isolation;
- b) Computer programmable amplifier;
- 15 c) Low-pass filter, with 1khz BandWidth;
- d) Clamp and Overvoltage protection to 5 V maximum;

MODULE 3: SWITCHING POWER SUPPLY

20

Circuit components of module 3 are listed:

- a) Fuse and Transformer;
- b) Diode Bridge and regulator ;
- c) Switching Power Supply +5v, 1A ;
- 25 d) Switching Power Supply +15v, 200mA;
- e) Switching Power Supply -15v, 200mA;

MODULE 4: BATTERY CHARGER PLUS BATTERY BACKUP SYSTEM

30

Circuit components of module 4 are listed:

- a) Battery charger ;
- b) Fuse ;
- c) Batteries ;

35

MODULE 5: GRAPHIC DIGITAL DISPLAY

The Graphic Digital Display is a component that integrates screen memory and controller, as well as all
40 necessary elements in order to interconnect with module 1, by means of a TTL compatible digital signal
with input-output ports. The communication protocol is indicated in program instructions herein contained.
There are many displays available that can be used. As an example, the VIKAY graphic series is proposed.

MODULE 6: KEYBOARD

45

Up to 16 key keyboard can be used, with either numerical and function keys. The keyboard is placed in
the font pannel, together with 2 visual indicators, one for Power ON/OFF state, the second as an alarm
indicator.

50

55

EQUIPMENT CONTROLS		
ITEM	NAME	FUNCTION
1	POWER Cord Connector	AC power cord connector with retaining mechanism.
2	DATA	For connection to external INTERFACE devices, such as DINAMAP or FINAPRES, or to an external computer. Screw-type connectors for connection of pump (for example NEW PERFUSOR SECURA of B.Braun). Controls the AC power to the monitor, plus the battery supply. Screw-type connector for connexion of pressure transducer (for example, COMBITRANS, of B.BRAUN)
3	PUMP CONNECTOR	
4	POWER ON-OFF	
5	PRESSURE TRANSDUCER CONNECTOR	

MOST SIGNIFICANT OPERATING FEATURES OF THE EQUIPMENT

► GENERAL FEATURES OF THE EQUIPMENT

The EQUIPMENT (arterial blood pressure monitor and controller) has been designed as a small portable bedside system.
The EQUIPMENT has been designed to provide operator convenience and reliability when monitoring blood pressure, and when acting as blood pressure supervisor, sending computer commands to the infusion pump in order to set the rate of drug to be delivered.

► DIFFERENT SITUATIONS WHERE THE EQUIPMENT CAN BE USED

As blood pressure monitor, the equipment finds the medical applications already widely documented in the specialized bibliography.
As pressure supervisor and controller, this EQUIPMENT is indicated in Operating Room and Medical Intensive Care Unit treatment, in the cases where a strict control of arterial pressure level is mandatory. Hypertensive crisis, acute myocardial infarction, congestive cardiac failure, cardiovascular pre, per and post-operative cardiac surgery, eclampsia, medical care (nephrology) are some examples of Intensive Care Unit applications. Controlled hypotension in neural surgery, otorrinolaryngology and plastic surgery (oscillometric non-invasive pressure measurement system) are some applications of Operating Room situations.
Under manual control, the EQUIPMENT permits the infusion of any drug, so that its use is recommended whenever it is advisable the infusion of a predefined drug, and the registration in memory of data such as blood pressure and dosage infused, on a minute by minute basis.

► MICROPROCESSOR BASED DESIGN.

System fully upgradable.

► BACK-UP BATTERY

The pressure supervisor is battery-backed and can be used as portable monitor during patient transportation.

► CONNECTION TO INFUSION PUMP

It can be connected via one cable to the NEW INFUSOR SECURA pump. The EQUIPMENT will feed the pump, (whenever there is a DC 3-20 volts input power connector available in the PUMP), and send computer based infusion commands in order to set the pump's infusion values. The requirements that pumps must fulfill are that they have a standard RS232-c or IEEE EIA485 (ELECTRONICS INDUSTRY ASOCIATION) computer interface connector. Due to the large available number of pumps of this kind in the medical field, it is imposible to refer to all of them. As an example, it can be used the B.Braun Infusor NEW PERFUSOR SECURA.

10 ► MEAN ARTERIAL PRESSURE MEASUREMENT SOURCES

While in monitor or control mode, pressure values can be taken from three different sources:

- 1) Intraarterial measurement by means of a percutaneous catheter. 2) Non-invasive oscilometric system (for example, DINAMAP of Critikon) and 3) Non-invasive pletismographic system (for example, FINAPRES of Omeda).

► The EQUIPMENT AS PRESSURE MONITOR

The EQUIPMENT can act as a monitor of blood pressure, installing the pressure transducer with a percutaneous intravenous catheter. the EQUIPMENT itself is actually designed with its own blood pressure monitor, making it unnecessary the use of an external monitor. It contains all the necessary elements in order to adapt the pressure transducer input signal. There is a large number of available transducers. The only requisite they must fulfill is that they be excited with DC supply voltage ranging between 2 to 15 volts. As an example, it is indicated the B.BRAUN COMBITRANS pressure transducer.

► CONNECTION TO AN EXTERNAL BLOOD PRESSURE MONITOR.

It can be also connected, besides its own monitoring system, to any external arterial blood pressure monitor (available in the medical field) containing an analog output signal in the range 0-5volts.

It can be also connected to different non-invasive blood pressure monitors, such as Oscilometric and Pletysmographic based systems, provided they have a standard RS232-c or IEEE EIA485 (ELECTRONICS INDUSTRY ASOCIATION) computer interface connector. As an example, it can be used DINAMAP (from Critikon) and FINAPRES (Omeda).

► INTERFACE RS232-c OR IEEE EIA485 (ELECTRONICS INDUSTRY ASOCIATION)

They are two Industry standard interface for compatibility with most external programmable devices and computer systems. This connector is used for connection to predefined non-invasive pressure monitor and computers such as for example IBM compatible computers.

45 ► ARTIFACT REJECTION.

The pressure supervisor is capable of eliminating most noise and motion artifact (monitor and control modes).

50 ► AUDIBLE/VISUAL ALARM SYSTEM

Provides a visual and audible indication if cardiovascular measurements fall outside of operator-programmable high/low limits and of abnormal system conditions or hardware failure.

55

► GRAPHIC DIGITAL DISPLAY

Easy to read. Any such DISPLAY can be used, provided it accepts communication with microprocessor board via input-output digital pots. Power supply can be in the range of 3 to 15 volts. As an example, VIKAY graphic display is indicated.

5

► UNITS OF MEASUREMENT

pressure is given in mmHg for pressure, and pulse in beats per minute (bpm).

10

MONITOR MODE:

► GRAPHIC DISPLAY

15

Monitors arterial pressure by means of a percutaneous implanted catheter. Results are graphically displayed on easy-to-read graphical display.

20

► PARAMETERS CALCULATION

Systolic and Diastolic are searched by a tracking algorithm. Mean Pressure and Pulse computation are performed.

25

CONTROL MODE:

► MANUAL AND AUTO CONTROL MODE

30

Two different control modes can be presented:

- 1) Manual: Any kind of drug can be used.
- 2) Auto: with Sodium-Nitroprusside only.

35

► SURVEILLANCE MODE

It can work in surveillance mode, starting infusion only when MAP is higher than a predefined value.

40

► PRESSURE CLAMP DETECTOR

It works as a pressure clamp detector. When clamp conditions are encountered, an acoustic alarm is activated.

45

► PUMP PROTOCOL

While in control mode, the EQUIPMENT will keep a continual communication protocol with the pump.
50 Any failure to keep the connection will be detected within seconds.

► DATA PRESENTATION

55 While in control mode, results of the last 30 minutes of each control are displayed on a graphic screen. Mean, Maximum and Medium Pressure are presented.

► DATA REGISTRATION

While in control mode, results of dose and pressure values are registered on a minute by minute basis. This values can be later reviewed both on the screen, or sent to a computer for further analysis. The
5 EQUIPMENT allows for 24 hour data registration.

► MANUAL MODE CONTROL

10 In manual control mode dosage is manually determined from the Baricontrol unit. Any drug can be used. Dosage will be maintained in this value until a new infusion command is given.

► AUTO MODE CONTROL

15 In auto control mode, only Sodium Nitroprusside must be used. Dosage are updated every 20 seconds according with a Fuzzy - Rule based algorithm for the computation of optimal dosage.

20 AUTOMATIC DOSAGE DETERMINATION

► INTELLIGENT COMPUTATION

25 The pressure supervisor carries out a sophisticated scheme for optimal dosage computation.

► ALGORITHM PRESENTATION

30 The algorithm introduces a fuzzy logic and rule-based procedure as a means of capturing human expertise and dealing with uncertainty. The algorithm incorporates "Fuzzy" logic elements, combined with an analytical optimal search methodology, for optimal dose computation.

35 ► PRESSURE TARGET

The ultimate objective of the algorithm is to bring pressure values to the target pressure gap. Target gap oscilates between target pressure and 10 mmHg above.

40 ► EASY TO UNDERSTAND

At all times, Dosage updating is very reasonable as a consequence of the computation method.

45 ► SMOOTH CONTROL

Optimal Dose search is carried out continually. There is however, minimal over-excitation to the patient.

50 ► MINIMAL DOSE INFUSION

In order to achieve a predefined target, the EQUIPMENT will adapt permanently, in order to reach it with minimum overall dosage infused to the patient. The method results minimally aggressive.
55

► COMPUTATION SCHEME OF OPTIMAL DOSE

The algorithm introduces a fuzzy logic and rule-based procedure as a means of capturing human expertise and dealing with uncertainty. The algorithm incorporates "Fuzzy" logic elements, combined with an analytical optimal search methodology, for optimal dose computation.

5

► CONTROL MECHANISM

Every 20 seconds, a new dose infusion must be determined. For this purpose, the MAP value is obtained, from over 1000 measures taken during a 5 seconds period.

10

► MEAN ARTERIAL PRESSURE CATEGORIZATION

Definition of a Target Pressure Value is done at the start of the control. The immediate consequence is that Pressure values become categorized. From a given value of Blood Pressure, four distinct areas are developed:

"Above Target Gap" Above Target Gap upwards.

"Target Gap", 5 mmHg wide above target pressure.

Whenever Pressure values remain in this gap, control is considered to have no error.

20 "Below Target Gap", about 5 mmHg wide.

"Risk Gap", 5mmHg below Pressure Target down wards. In this case, Dose infusion is always 0.

► OPTIMAL DOSE CONCEPT

25

The ultimate objective of the algorithm is to bring pressure values to the target gap. Once the target gap is reached, the Optimal Dose is infused, in order to maintain pressure values within its bounds. The Optimal Dose value is a theoretical concept, by which we mean a dose that, continually infused, will maintain pressure values within the target gap. Of course, this value adaptively changes with time and with varying patient conditions, and a permanent updating is carried out, whenever the pressure evolution undergoes an excursion either above or below the target gap.

30

► DOSE DETERMINATION

35

Two different steps are conducted in order to determine the new infusion value.

► FIRST STEP IN DOSE DETERMINATION

40

First step is conducted following the fuzzy logic methodology. The objective is to determine what kind of action should be taken. This decision has only three possible options: to maintain, to increase or to decrement dose infusion values. The information required to perform this step is based on qualitative rather than quantitative values of variables. Linguistic labels are associated to the values obtained of input variables such as last MAP sampled, trajectory followed by pressure (shape factor, mean slope value, etc) and pressure gap as well as proximity to target gap.

45

► DATA CATEGORIZATION

50

These labels are then related through conditional declarations in order to categorize the information. Priority decisions will determine the kind of action to be taken whenever the labels yield a vague or undefined situation. Once this first step is carried out, it remains still the need to determine the quantitative value of new dose to be infused.

55

► SECOND STEP IN DOSE DETERMINATION

The second step implements an optimal dose search algorithm based on a Rule-Based methodology. The objective is twofold: On one side, The new dose value must be determined. If MAP is not in the target gap, then new Infusion values are incremented or decremented relative to the actual dose value, through a non linear relationship with actual MAP. The Proportionality coefficient increases exponentially with MAP. On
5 the other side, an updating of optimal dose value is performed. If MAP belongs to the target gap, then the new infusion value equals the optimal dose. Otherwise, optimal dose is increased or decreased more slowly, following a logarithmic incrementation coefficient. This logarithmic relationship adapts with the patient's varying situation by means of a growth factor.

10

► NOTE:

The computer programs have been appended, in case some details may have escaped from the
15 present indications.

20

25

30

35

40

45

50

55

```

/*
 *   PROGRAM INSTRUCTIONS.
5  *   Authors:  Dr.Daniel Borches
 *               Dr. Ricardo Ruiz
 */
10 unsigned char punto[1664]; /* 16*105 puntos al monitorizar */
unsigned short dosis_sim[90]={
15     356,356,356,356,356,356,356,356,382,412,
     440,468,710,466,0,0,0,0,0,148,
     256,256,256,256,128,128,0,0,104,184,
20     184,184,184,184,184,92,92,92,82,84,
     148,148,148,148,148,148,74,74,74,74,
     116,116,116,116,116,116,116,116,116,58,
25     58,58,104,104,104,104,104,104,104,104,
     104,104,104,104,104,104,104,104,104,104,
     104,104,104,104,104,104,104,104,104,104
30 };

unsigned char presion_sim[90]={
35     89,89,90,89,90,90,90,90,91,91,
     91,91,89,77,63,68,65,65,63,65,
     74,76,74,71,69,66,64,64,67,71,
40     74,75,74,71,70,68,65,65,67,68,
     70,72,72,72,71,70,69,68,67,70,
     71,72,74,73,72,72,71,70,71,69,
45     69,70,71,73,73,73,73,72,72,72,
     71,72,72,70,71,70,70,71,71,71,
     71,71,71,71,71,71,71,71,71,71
50 };

55

```

```

unsigned char grafico_sim[110]={
5
    20,20,19,19,18,18,18,18,17,17,
    17,17,16,16,15,15,15,15,14,14,
    13,13,18,18,21,21,25,25,27,27,
10
    30,30,28,28,25,25,23,23,21,21,
    20,20,19,19,18,18,18,18,17,17,
    17,17,16,16,15,15,15,15,14,14,
    13,13,18,18,21,21,25,25,27,27,
15
    30,30,28,28,25,25,23,23,21,21,
    20,20,19,19,18,18,18,18,17,17,
    17,17,16,16,15,15,15,15,14,14,
20
    13,13,18,18,21,21,25,25,27,27,
    };

25
unsigned char fin_programa;
#include "stdio.h"
/*
30
    *   Declaracion de prototipos.
    */
void control_automatico(void);
35
void control_manual(void);
void monitorizar_presion(void);
void finalizar_programa(void);
40
void opciones_varias(void);

45

50

55

```

```

/*
5   *   Funciones.
   */

void dibuja_presion_dosis(coord_x,presion,ant_dosis,dosis)
10  unsigned char coord_x;
    unsigned char presion;
    unsigned short ant_dosis;
15  unsigned short dosis;
    {
        unsigned char j,p=0,a_d=0,d=0;
20  unsigned short j2;
        if(coord_x == 1) ant_dosis = dosis;
        j=presion;while(j>=5){j-=4;p++;}
25  j2=ant_dosis;while(j2>=21){j2-=20;a_d++;}
        j2=dosis;while(j2>=21){j2-=20;d++;}
        for (j=0;j<=p;j++)plot(71-j,coord_x+23);
30  if(d>p){
            plot(71-d,coord_x+23);
            if (a_d<=d)
35  {
                if (a_d >= p)
                    for (j=(a_d+1);j<d;j++) plot(71-j,coord_x+23);
40  else
                    {
                        for (j=(a_d+1);j<p;j++) unplot(71-j,coord_x+23);
45  for (j=(p+1);j<d;j++) plot(71-j,coord_x+23);
                    }
                }
50
55

```

```

        else
5           for (j=(d+1);j<a_d;j++) plot(71-j,coord_x+23);
        }
    else {
10       unplot(71-d,coord_x+23);
        if (a_d <= d)
            for (j=(a_d+1);j<d;j++) unplot(71-j,coord_x+23);
15       else
        {
            if (a_d<=p)
20         for (j=(d+1);j<a_d;j++) unplot(71-j,coord_x+23);
            else
            {
25         for (j=(d+1);j<p;j++) unplot(71-j,coord_x+23);
            for (j=(p+1);j<a_d;j++) plot(71-j,coord_x+23);
            }
30         }
        }
    }

void control_de_presion( modo )
35 char modo; /* modo=0 : control automatico */
            /* modo=1 : control manual      */

40 {
    extern struct param valor_inicial;
    extern char hay_alarma;
    extern char l_pressed, kb_nextin;
45    extern int t_20s;

```

50

55

```

    unsigned short dosis[90];
    unsigned char fin_control=0,fin_bloque=0,i=1,j,k,descenso_lento=0
5      forma,simulado=0,
      presion_final=70,presion_media,
      presion_maxima=0,presion_minima=0xff, a_d,
10     presion_alarma=200 ,presion[90],
      consulta,i_cont=0;

    unsigned short total,dosis_ant=0,dosis_auxiliar,
15     dosis_inicial=50 /* 0.5ml/h */;

    unsigned long p_sum;

    short dosis_actual;
20   char fallo;

    dosis_actual=dosis_inicial;
25   forma=modo;

    while (!fin_control){
        kb_nextin=0;
30     fin_bloque=0;
        hay_alarma = 0;

        while(!fin_bloque && !hay_alarma){
35             consulta = 0;
            borra_grafico();
            if (modo) {
40                 menu_4_v("CONTRL","MENU","Real","Simul");
                display_string_h(7,2,"Control:");
                if (!simulado) display_string_h(7,10,"REAL    ");
45                 else      display_string_h(7,10,"SIMULADO");
            }

            else {
50
55

```



```

        menu_4_v("CONTRL","MENU","D.rap","D.len");
        display_string_h(7,2,"Descenso :");
5         if (descenso_lento) display_string_h(7,12,"LENTO ");
            else             display_string_h(7,12,"RAPIDO");
        }

10

        display_string_h(1,2,"0 Dosis I:");
        display_var(1,16,dosis_actual);
15        display_string_h(3,2,"1 P.Final:");
        display_char(3,15,presion_final);
        display_string_h(5,2,"2 P.Alarm:");
20        display_char(5,15,presion_alarma);

        kb_pressed=0;
25        while(!kb_pressed);
        switch (kb_pressed){
            case 'A':  fin_bloque=1;
30                     consulta = 1;
                     borra_grafico();
                     menu_4_v("DOSIS","MENU","", "");
35                     display_string_v(0,19,"");
                     display_string_v(0,18,"");
                     dibuja_ejes();
40                     for (k=1;k<=i;k++){
                        dibuja_presion_dosis(k,presion[k],dosis[k-1],dosis[k]);
                        }
45                     inicializa_conversion();
                     break; /* Opcion sup. izda. */

50

55

```

```

case 'B': fin_bloque=1;
          fin_control=1;
5          break; /* Opcion inf. izda. */
case 'C': if (modo) simulado=0;
          else descenso_lento=0;
10          break; /* Opcion sup. dcha. */
case 'D': if (modo) simulado=1;
          else descenso_lento=1;
15          break; /* Opcion inf. dcha. */
case '0': borra_grafico();
          display_string_h(3,5,"Introduzca");
20          display_string_h(4,9,"la");
          display_string_h(5,4,"DOSIS INICIAL");
          display_string_h(6,6,"( ml./h. )");
25          dosis_inicial=input_var(8,10,dosis_inicial);
          dosis_actual=dosis_inicial;
          break; /* Opcion 0 */
30 case '1': borra_grafico();
          display_string_h(3,5,"Introduzca");
          display_string_h(4,9,"la");
35          display_string_h(5,4,"PRESION FINAL");
          display_string_h(6,6,"( mm.Hg. )");
          presion_final=input_char(8,10,presion_final);
40          break; /* Opcion 1 */
case '2': borra_grafico();
          display_string_h(3,5,"Introduzca");
45          display_string_h(4,9,"la");
          display_string_h(5,2,"PRESION ALARMA");
          display_string_h(6,6,"( mm.Hg. )");
50
55

```

```

                                presion_alarma=input_char(8,10,presion_alarma)
5                                break; /* Opcion 1 */
                                default : break;
                                }
10        }
    if (consulta == 1)
    {
15        datos_perfusor(&fallo);
        alarmas_perfusor(&fallo);

20        valor_inicial.presion_objetivo = presion_final;
        dosis_auxiliar = dosis_actual/10;
        valor_inicial.infusion_inicial = dosis_auxiliar;
25        if (descenso_lento)    valor_inicial.ritmo_caida = 10;
            else                valor_inicial.ritmo_caida = 15;

30

        p_sum=0;total=0;
        t_20s=155;
35        while (t_20s){
            inicializa_adctl(0,0,0,0,0,0);
            while( !fin_conversion() );
40            a_d=*adrl_addr;
            p_sum=p_sum+a_d;
            total++;
45        }

50

55

```

```

presion_media = 0;
    while (p_sum>total){p_sum=p_sum-total;presion_media++; }
5      valor_inicial.presion_inicial = presion_media;
      inicia_control(valor_inicial);
    }
10  fin_bloque=0;
    l_pressed=0;
    while( (!fin_bloque) && (!fin_control) && (!hay_alarma)){
15      t_20s=400;
      while (t_20s);
      p_sum=0;total=0;
20      l_pressed=0;
      datos_perfusor(&fallo);
      if (hay_alarma) break;
25      alarmas_perfusor(&fallo);
      if (hay_alarma) break;
      t_20s=155;
30      while (t_20s){
          inicializa_adctl(0,0,0,0,0,0);
          while( !fin_conversion() );
35          a_d=*adrl_addr;

          if(!total) presion_minima = presion_maxima = a_d;
40          if(a_d>presion_maxima) presion_maxima=a_d;
          if(a_d<presion_minima) presion_minima=a_d;

```

45

50

55

```

    p_sum=p_sum+a_d;
    total++;
5    if (l_pressed=='A'){
        forma=forma+2;
        l_pressed=0;
10    display_string_h (1,13,"      ");
        display_string_h (2,13,"      ");
        display_string_h (3,13,"      ");
15    display_string_h (4,13,"      ");
        display_string_h (5,13,"      ");
        display_string_h (6,13,"      ");
20    dosis_actual=input_var(9,18,dosis_inicial);
        if (!modo)
        {
25            dosis_auxiliar = dosis_actual/10;
            valor_inicial.infusion_inicial = dosis_auxiliar;
            inicia_control(valor_inicial);
30        }
        t_20s=0;
    }
35 }

    presion_media=0;
    while (p_sum>total){p_sum=p_sum-total;presion_media++; }
40    if (l_pressed == 'B') break;
    if (simulado) presion_media=presion_sim[i];
    presion[i]=presion_media;
45    if (!forma){
        dosis_actual = dosis_control(presion[i]);
        dosis[i] = dosis_actual;
50
55

```

```

        }
        else {
5           dosis[i] = dosis_actual;
        }
10      if (simulado) { dosis[i]=dosis_sim[i];dosis_actual = dosis[i]; }

      if (forma>1) forma=forma-2;
15      if (dosis[i] < 0.1) dosis[i] = 0;
      dibuja_presion_dosis(i,presion[i],dosis[i-1],dosis[i]);

20      if((dosis_actual == 0) && (dosis_ant == 0)) i_cont++;
      else i_cont = 0;
      if (i_cont == 4)
25      {
          i_cont = 0;
          dosis_actual = 1;
30      }
      t_20s = 31;
      while(t_20s);
35      ajuste_perfusor(dosis_actual,5000,&fallo);
      t_20s = 3;
      while(t_20s);
40      i++;
      if (i==90){ /* scroll horizontal */
          i=45;
45      borra_grafico();
          menu_4_v("DOSIS","MENU","", "");
          display_string_v(0,18,"
50

```

55

```

        display_string_v(0,19,"          ");
5      dibuja_ejes();
        dosis[0]=dosis[45];
        presion[0]=presion[45];
10     for (k=1;k<=44;k++){
            dosis[k]=dosis[k+45];
            presion[k]=presion[k+45];
15     dibuja_presion_dosis(k,presion[k],dosis[k-1],dosis[k]);
        }
    }

20  display_string_h (0,16,"mmHg");
    display_string_h (1,17,"med");
    display_char (2,19,presion_media);
25  display_string_h (3,17,"max");
    display_char (4,19,presion_maxima);
    display_string_h (5,17,"min");
30  display_char (6,19,presion_minima);
    display_string_h (8,17,"INF");
    display_var (9,18,dosis_actual);
35  switch (l_pressed){
        case 'A':  forma=forma+2;
                    display_string_h (0,16,"          ");
                    display_string_h (1,17,"          ");
40                    display_string_h (2,17,"          ");
                    display_string_h (3,17,"          ");
                    display_string_h (4,17,"          ");
45                    display_string_h (5,17,"          ");
                    display_string_h (6,17,"          ");
                    dosis_actual=input_var(9,18,dosis_inicial);
50

```

55

```

        if (!modo)
5           {
                dosis_auxiliar = dosis_actual/10;
                valor_inicial.infusion_inicial = dosis_auxiliar;
10                inicia_control(valor_inicial);
                }

                break; /* Opcion sup. izda. */

15        case 'B': fin_bloque=1;
                break; /* Opcion inf. izda. */

                default : break;
20        }

        l_pressed=0;
        dosis_ant = dosis_actual;
25        }
    }

    finaliza_conversion();
    ajuste_perfusor(0,0,&fallo);
30    }

    void control_automatico(void)
    {
35        control_de_presion(0);
    }

40    void control_manual(void)

        control_de_presion(1);
45    }

```

50

55


```

void monitorizar_presion(void)
{
5     extern char l_pressed;
    extern unsigned char punto[1664], grafico_sim[110];
    char fin_monitorizacion=0;
10    unsigned char i,j,k,num_punto,t=16,simulado,
        b,x_byte,y_byte,minimo=180,
        presion_media,presion_maxima=0,presion_minima=0xff;
15    unsigned short total;
    unsigned long p_sum;

20    borra_grafico();

    menu_4_grande("", "",
25        "SIMULADA", "",
        "", "",
        " REAL", "");

30    display_string_h(1,0,"");
    display_string_h(2,0,"");
    display_string_h(3,0," MONITORIZACION ");
35    display_string_h(4,0,"");
    j=1;
    while(j){
40        l_pressed=0;
        while (!l_pressed);
        switch (l_pressed){
45            case 'B': simulado=1;
                j=0;
                break;
50
55

```

```

        case 'D': simulado=0;
                    j=0;
                    break;
        default : break;
    }
}

borra_grafico();
menu_4_v("STOP ", "MENU", "Per.+", "Per.-");
inicializa_conversion();
l_pressed=0;
while(!fin_monitorizacion){
    y_byte=180;
    total=0;
    while((y_byte>(minimo+2))&&(total<2000)){
        *adctl_addr=0; /* SCAN=0 MULT=1      ANO */
        while(!fin_conversion());
        y_byte=(*adrl_addr);
        y_byte=y_byte/4;
        y_byte=71-y_byte;
        total++;
    }

    minimo=180;
    p_sum=0;
    total=0;
    *adctl_addr=0;

    num_punto=0;
    if (simulado) t=16;

```

```

for (i=2;i<=14;i++) {
    for (b=0;b<=7;b++) {
5         for (k=0;k<=t-1;k++) {
            x_byte=i;
            y_byte=*adrl_addr;
10         *adctl_addr=0;

            if(!k) presion_minima = presion_maxima = y_byte;
            if(y_byte<presion_minima) presion_minima=y_byte;
15         if(y_byte>presion_maxima) presion_maxima=y_byte;
            p_sum+=y_byte;
            y_byte=y_byte/4;
20         y_byte=71-y_byte;
            if ((y_byte>80)) y_byte=80;
            if (simulado) y_byte= grafico_sim [num_punto];
25         if (y_byte<minimo) minimo=y_byte;
            punto[(short)(num_punto+(110*k))]=y_byte;
                /* Almacena valor actual */
30         if (i<14)
*yx_display=(short)(i+1+20*punto[(short)(8+num_punto+(110*k))]);

                                                    e l s e
35         *yx_display=(short)(2+20*punto[(short)(b+(110*k))]);
            *portb_addr=32 | *portb_addr;
            *vis_addr=0x0a;
40         *portb_addr=128 | *portb_addr;
            *portb_addr=127 & *portb_addr;
            *portb_addr=223 & *portb_addr;
45         *vis_addr=*x_display;
            *portb_addr=128 | *portb_addr;
            *portb_addr=127 & *portb_addr;
50
55

```

```

5      *portb_addr=32 | *portb_addr;
      *vis_addr=0x0b;
      *portb_addr= 128 | *portb_addr;
      *portb_addr=127 & *portb_addr;
10     *portb_addr=223 & *portb_addr;
      *vis_addr=*y_display;
      *portb_addr=128 | *portb_addr;
15     *portb_addr=127 & *portb_addr;
      *portb_addr=32 | *portb_addr;
      *vis_addr=0x0e;
20     *portb_addr=128 | *portb_addr;
      *portb_addr=127 & *portb_addr;
      *portb_addr=223 & *portb_addr;
25     *vis_addr=b;
      *portb_addr=128 | *portb_addr;
      *portb_addr=127 & *portb_addr;
30
      *yx_display=(short)(x_byte+20*y_byte);

35     *portb_addr=32 | *portb_addr;
      *vis_addr=0x0a;
      *portb_addr=128 | *portb_addr;
40     *portb_addr=127 & *portb_addr;
      *portb_addr=223 & *portb_addr;
      *vis_addr=*x_display;
45     *portb_addr=128 | *portb_addr;
      *portb_addr=127 & *portb_addr;
      *portb_addr=32 | *portb_addr;
50     *vis_addr=0x0b;

```

55

```

        *portb_addr= 128 | *portb_addr;
        *portb_addr=127 & *portb_addr;
5      *portb_addr=223 & *portb_addr;
        *vis_addr=*y_display;
        *portb_addr=128 | *portb_addr;
10     *portb_addr=127 & *portb_addr;
        *portb_addr=32 | *portb_addr;
        *vis_addr=0x0f;
15     *portb_addr=128 | *portb_addr;
        *portb_addr=127 & *portb_addr;
        *portb_addr=223 & *portb_addr;
20     *vis_addr=b;
        *portb_addr=128 | *portb_addr;
        *portb_addr=127 & *portb_addr;
25     total++;
    }
    num_punto++;
30 }
}

    num_punto--;
35     presion_media=0;
    while (p_sum>total){p_sum=p_sum-total;presion_media++;}
    display_string_h (0,15,"mmHg");
40     display_string_h (2,15,"med");
    display_char (3,17,presion_media);
    display_string_h (5,15,"max");
45     display_char (6,17,presion_maxima);
    display_string_h (8,15,"min");
    display_char (9,17,presion_minima);
50
55

```

```

switch (l_pressed){
5      case 'A':  menu_4_v("SIGUE","      ","      ","      ");
              l_pressed=0;
              while (l_pressed!='A');
10      menu_4_v("STOP","MENU","Per.+","Per.-");
              break; /* Opcion sup. izda. */
      case 'B':  fin_monitorizacion=1;
15      break; /* Opcion inf. izda. */
      case 'C':  if (t<16) {
                  t=t*2;
20      borra_grafico();
                  menu_4_v("STOP","MENU","Per.+","Per.-");

25      }
              break; /* Opcion sup. dcha. */
      case 'D':  if (t>1) {
                  t=t/2;
30      borra_grafico();
                  menu_4_v("STOP","MENU","Per.+","Per.-");

35      }

40      break; /* Opcion inf. dcha. */
      default :  break;
              }
45      l_pressed=0;
              }
      finaliza_conversion(); /* Apaga el conversor AD */
50
55

```

```

    }

5

void finalizar_programa(void)
10 {
    extern char l_pressed;

15    borra_grafico();

    menu_4_grande("", "",
20        " SI", "",
        "", "",
        " NO", "");

25    display_string_h(1,0,"      F I N      ");
    display_string_h(2,0,"      del      ");
    display_string_h(3,0,"    Control de    ");
30    display_string_h(4,0," Presion Arterial ?");

35    while(1){
        l_pressed=0;
        while (!l_pressed);
40        switch (l_pressed){
            case 'B': fin_programa=1;
                    borra_grafico();
55                    display_string_h(2,1,"Programa terminado");
                    break;
            case 'D': fin_programa=0;
50                    break;
        }
    }

```

```

        default : break;
5          }
    }
}

10
void opciones_varias(void)
{
15     extern char l_pressed, kb_pressed;
    char fin_bucle;
    fin_bucle = 0;
20     while(!fin_bucle){
        borra_grafico();      /* Vacía la pantalla */

25     menu_4_grande("  LUZ   ", "PANTALLA",
                    "  FIN   ", "PROGRAMA",
                    "  MENU  ", "PRINCIPAL",
30     "  SALIDA ", "ARCHIVOS");
        l_pressed=0;
        while (!l_pressed);    /* Mientras no se elija una opción */
35     switch (l_pressed){
        case 'A': borra_grafico();
                    display_string_h(3,2,"0 ILUMINAR FONDO");
40                    display_string_h(5,2,"1 APAGAR FONDO");
                    kb_pressed = 0;
                    while(!kb_pressed);
45                    switch (kb_pressed) {
                        case '0': reset_bit (1,portb_addr);
                                    break;
50                        case '1': set_bit (1,portb_addr);

```



```

                                break;
                                default : break;
5                                }

                                break;

10                                case 'B': finalizar_programa();
                                break;

                                case 'C': fin_bucle = 1;
15                                break;

                                case 'D': break;
                                default : break;
20                                } -
                                }

25                                }
main()
{
30    extern char l_pressed;
    char fin_programa=0;
    *init_addr = 0x01;
35    *pioc_addr = 0x43;
    *ddrc_addr = 0xff;
40    inicializa_irq();
    intercept_serial_6811();
    inicializa_serie_6811();
45    inicializa_serie_68681();
    inicializa_newtimer();
50    modo_grafico();
    borra_grafico();

55    menu_1();

```

```

t_20s = 100;
while(t_20s);

5

while(!fin_programa){
10    borra_grafico();
    menu_4_grande(" Control","AUTOMAT.",
                  " Control"," MANUAL ",
15    " MONITOR","          ",
                  "OPCIONES"," VARIAS ");

20

    l_pressed=0;
    while (!l_pressed);
25    switch (l_pressed){
        case 'A': control_automatico();
            break; /* Opcion sup. izda. */
30    case 'B': control_manual();
            break; /* Opcion inf. izda. */
35    case 'C': monitorizar_presion();
            break; /* Opcion sup. dcha. */
40    case 'D': opciones_varias();
            break; /* Opcion inf. dcha. */
        default : break;
45    }
    }

50

while(1);

55 }

```

Claims

5

1. Improvements to methods of delivering vasoactive drugs by suitable infusion devices, controlled by a computerized device which controls and adjusts dosage automatically in a closed loop and is used to control blood pressure. These improvements consist of functionally associating a controller (1), a monitor (2) and an infusion pump (3). The controller (1) includes a micro-controller, ROM memory, RAM memory,

10

visualizer, three RS-232 series communication outputs, real time clock, keyboard, indicators, microprocessor system backup logic, battery, battery charging system, battery charging level detector and power supply board.

2. Improvements to methods of delivering vasoactive drugs by suitable infusion devices, using a computerized device which controls and adjusts dosage automatically in a closed loop and is used to control blood

15

pressure, as per claim 1, which introduces a new concept in control based on qualitative characterization of the variables to be controlled. This controller applies logical analysis to describe these "very complex" or "badly defined" systems which do not accept precise mathematical analysis.

3. Improvements to methods of delivering vasoactive drugs by infusion pumps using a computerized device which controls and adjusts dosage automatically in a closed loop and is used to control blood pressure, as

20

per above claims, whose main features are as follows:

3.1. The use of qualitative characterization (linguistic) of the control stage, rather than quantitative variables (numerical) and

3.2. The establishment of relationships between variables by means of conditional statements which include Bool's algebra rules.

25

4. Improvements to methods of delivering vasoactive drugs by infusion pumps using a computerized device which controls and adjusts dosage automatically in a closed loop and is used to control blood pressure, as per claim 1, which is characteristic because the microcontroller has a central processing unit, a series communication device, two parallel input/output communication, doors of eight bits each and eight analogical digital conversion channels, having on/off indicators, as well as battery level, running control and

30

printing order indicators.

5. Improvements to methods of delivering vasoactive drugs by infusion pumps using a computerized device which controls and adjusts dosage automatically in a closed loop and is used to control blood pressure, as per claim 1, which is characteristic because the microprocessor system backup logic has decodifying mechanisms, reset generation, arterial pressure reference and latch.

35


6. Improvements to methods of delivering vasoactive drugs by infusion pumps using a computerized device which controls and adjusts dosage automatically in a closed loop and is used to control blood pressure, as per the above claims, which is characteristic because the power supply board has a DC-DC converter, insulation transformer, rectifier, regulator and mains-battery switch.

40

45

50

55





European Patent
Office

EUROPEAN SEARCH REPORT

Application Number

EP 90 50 0063

DOCUMENTS CONSIDERED TO BE RELEVANT			
Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (Int. Cl.5)
X	US-A-4 392 849 (PETRE et al.) * Figure 1; column 1, lines 65-67; column 2, lines 6-53; column 5, lines 7-39; column 6, lines 26-41; claim 2 *	1,3,5	A 61 M 5/172
Y	---	2,6	
A	---	4	
X	US-A-4 551 133 (ZEGERS DE BEYL et al.) * Figures 1,6; column 2, lines 19-60; column 5, lines 41-56; column 5, line 67 - column 6, line 12; column 6, lines 17-32 *	1,4	
X	DE-A-2 741 344 (RUTTIGERS) * Figure 1; page 11, lines 17-21; page 12, line 16 - page 13, line 12; page 13, line 21 - page 14, line 3; page 18, line 13 - page 19, line 3 *	1,2	
Y	US-A-4 525 775 (EYDELMAN) * Column 7, lines 12-28; column 8, lines 11-28 *	2	TECHNICAL FIELDS SEARCHED (Int. Cl.5)
Y	EP-A-0 164 904 (THE UNIVERSITY OF MELBOURNE) * Figure 7; page 16, lines 8-11 * --- -/-	6	A 61 M
The present search report has been drawn up for all claims			
Place of search THE HAGUE		Date of completion of the search 12-10-1990	Examiner SEDY, R.
CATEGORY OF CITED DOCUMENTS			
X : particularly relevant if taken alone Y : particularly relevant if combined with another document of the same category A : technological background O : non-written disclosure P : intermediate document		T : theory or principle underlying the invention E : earlier patent document, but published on, or after the filing date D : document cited in the application L : document cited for other reasons & : member of the same patent family, corresponding document	



European Patent
Office

EUROPEAN SEARCH REPORT

Application Number

EP 90 50 0063

DOCUMENTS CONSIDERED TO BE RELEVANT			
Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (Int. Cl.5)
A	<p>ADVANCES IN INSTRUMENTATION, vol. 39, parts 1,2, PROCEEDINGS OF THE ISA INTERNATIONAL CONFERENCE AND EXHIBIT, Houston, Texas, 22nd - 25th October 1984, pages 1082-1093, ISA, Research Triangle Park, NC, US; MOORE et al.: "Expert systems applications in industry"</p> <p>* Figures 3,5; page 1081, lines 27-34,44-48; page 1084, lines 63,64; page 1085, lines 24-30 *</p> <p>-----</p>	2	
			TECHNICAL FIELDS SEARCHED (Int. Cl.5)
The present search report has been drawn up for all claims			
Place of search THE HAGUE		Date of completion of the search 12-10-1990	Examiner SEDY, R.
<p>CATEGORY OF CITED DOCUMENTS</p> <p>X : particularly relevant if taken alone Y : particularly relevant if combined with another document of the same category A : technological background O : non-written disclosure P : intermediate document</p> <p>T : theory or principle underlying the invention E : earlier patent document, but published on, or after the filing date D : document cited in the application L : document cited for other reasons</p> <p>..... & : member of the same patent family, corresponding document</p>			